
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000831
11:04 31/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:16
PRESSURIZING 5:55
STERILIZING 7:20
136.0 C 220kPa 7:20
136.4 C 223kPa 7:50
136.0 C 220kPa 8:20
136.4 C 221kPa 8:50
136.1 C 220kPa 9:20
136.0 C 220kPa 9:50
136.4 C 222kPa 10:20
136.2 C 222kPa 10:50
136.2 C 222kPa 10:50
VENTING 10:50
AIR DRYING 11:51
CYCLE COMPLETE 15:49

Digital Signature #
8C1CCE2B58BA585F
