
Midmark M11 - v1.0.5
Total Cycles: 2995

Name: RITTER2
Sterilizer ID: V2313228
10/11/2023 10:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.1 F 0.0
2:00 186.7 F 2.9
4:00 200.0 F 8.1
6:00 217.8 F 10.2
8:00 230.8 F 10.1
10:00 233.0 F 8.0
12:00 240.3 F 10.8
14:00 247.4 F 14.4
16:00 260.3 F 21.6
18:00 272.0 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 273.0 F 29.7
2:00 271.5 F 29.0
3:00 273.1 F 29.8
4:00 272.1 F 29.2

Min 271.4 F 28.9
Max 273.2 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.2 F 0.0
5:00 230.2 F 0.0
10:00 232.0 F 0.0
15:00 227.5 F 0.0
20:00 224.2 F 0.0
25:00 224.6 F 0.0
30:00 221.0 F 0.0
Drying Complete

FILLING: 2:16
HEATING: 18:32
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:58
CYCLE COMPLETE

10/11/2023 11:37 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____