
Midmark M11 - v1.0.0
Total Cycles: 2004

Name: RITTER 1
Sterilizer ID: V2178198
06/07/2022 02:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.7 F 0.0
2:00 195.0 F 4.1
4:00 213.0 F 9.1
6:00 227.3 F 6.4
8:00 233.6 F 10.2
10:00 236.7 F 7.6
12:00 244.5 F 13.4
14:00 243.1 F 12.1
16:00 256.2 F 18.8
18:00 267.5 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.5 F 29.3
2:00 271.5 F 28.5
3:00 272.0 F 29.2
4:00 272.5 F 29.1

Min 271.2 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.7 F 0.0
5:00 220.3 F 0.0
10:00 226.5 F 0.0
15:00 230.2 F 0.0
20:00 228.8 F 0.0
25:00 229.8 F 0.0
30:00 225.0 F 0.0

Drying Complete

FILLING: 1:59
HEATING: 19:23
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:36
CYCLE COMPLETE

06/07/2022 04:00 PM
ACCEPTED: X
REJECTED:
Operator: U02