
Midmark M11 - v1.0.5
Total Cycles: 2257

Name: RITTER2
Sterilizer ID: V2313228
04/14/2023 04:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.6 F 0.0
2:00 180.2 F 3.1
4:00 188.9 F 7.6
6:00 216.3 F 12.9
8:00 225.8 F 7.8
10:00 238.2 F 13.1
12:00 238.6 F 10.9
14:00 237.0 F 9.7
16:00 251.1 F 16.3
18:00 262.5 F 22.8
20:00 272.8 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.5 F 29.4
2:00 271.1 F 28.7
3:00 272.5 F 29.6
4:00 272.0 F 29.1

Min 271.1 F 28.7
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.8 F 0.0
5:00 213.6 F 0.0
10:00 224.0 F 0.0
15:00 215.4 F 0.0
20:00 211.3 F 0.0
25:00 211.1 F 0.0
30:00 211.6 F 0.0
Drying Complete

FILLING: 2:04
HEATING: 20:22
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 00:59:48
CYCLE COMPLETE

04/14/2023 05:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____