

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3111

Name: RITTER 1  
Sterilizer ID: V2178198  
04/14/2023 07:24 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 75.1 F 0.0  
2:00 77.8 F 0.0  
4:00 104.0 F 0.0  
6:00 120.7 F 0.0  
8:00 162.3 F 0.6  
10:00 174.1 F 3.8  
12:00 209.9 F 7.0  
14:00 203.7 F 10.6  
16:00 218.7 F 7.1  
18:00 226.5 F 7.6  
20:00 236.6 F 11.7  
22:00 232.7 F 7.9  
24:00 245.1 F 13.5  
26:00 243.0 F 11.8  
28:00 255.0 F 17.9  
30:00 265.6 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.5 F 28.9  
2:00 271.5 F 28.3  
3:00 272.4 F 28.9  
4:00 271.5 F 28.2

Min 271.5 F 28.2  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.3 F 0.0  
5:00 195.1 F 0.0  
10:00 210.6 F 0.0  
15:00 221.3 F 0.0  
20:00 225.6 F 0.0  
25:00 228.3 F 0.0  
30:00 222.7 F 0.0  
Drying Complete

FILLING: 0:47  
HEATING: 31:49  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:09:49  
CYCLE COMPLETE

04/14/2023 08:33 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_