
Midmark M11 - v1.0.5
Total Cycles: 1472

Name: RITTER2
Sterilizer ID: V2313228
09/12/2022 07:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.6 F 0.0
2:00 78.8 F 0.0
4:00 93.3 F 0.0
6:00 104.8 F 0.0
8:00 152.2 F 0.1
10:00 176.8 F 2.7
12:00 191.2 F 6.2
14:00 208.5 F 9.8
16:00 223.2 F 11.7
18:00 226.7 F 8.4
20:00 236.2 F 12.8
22:00 237.1 F 9.7
24:00 237.8 F 9.5
26:00 246.6 F 13.9
28:00 257.8 F 19.9
30:00 267.8 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.9
1:00 272.2 F 29.6
2:00 271.6 F 29.1
3:00 272.2 F 29.1
4:00 272.0 F 29.7

Min 271.4 F 28.8
Max 272.4 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.8 F 0.0
5:00 228.2 F 0.0
10:00 222.2 F 0.0
15:00 219.7 F 0.0
20:00 222.8 F 0.0
25:00 225.0 F 0.0
30:00 219.7 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 31:14
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 01:09:06
CYCLE COMPLETE

09/12/2022 08:35 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____