
Midmark M11 - v1.0.5
Total Cycles: 3197

Name: RITTER2
Sterilizer ID: V2313228
12/08/2023 11:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.6 F 0.0
2:00 136.2 F 0.5
4:00 140.6 F 2.6
6:00 165.8 F 5.8
8:00 188.7 F 9.7
10:00 204.1 F 9.0
12:00 221.4 F 7.7
14:00 234.3 F 12.4
16:00 233.4 F 9.1
18:00 239.1 F 10.5
20:00 245.2 F 13.3
22:00 256.7 F 19.5
24:00 267.3 F 26.1
26:00 273.0 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.4
1:00 272.8 F 29.8
2:00 271.5 F 29.6
3:00 272.8 F 29.8
4:00 271.5 F 29.4

Min 271.4 F 29.3
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.3 F 0.0
5:00 215.8 F 0.0
10:00 219.1 F 0.0
15:00 210.4 F 0.0
20:00 206.1 F 0.0
25:00 207.7 F 0.0
30:00 210.4 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 26:20
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:04:47
CYCLE COMPLETE

12/08/2023 12:45 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____