
Midmark M11 - v1.0.5
Total Cycles: 112

Name: RITTER2
Sterilizer ID: V2313228
06/17/2021 10:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.7 F 0.0
2:00 157.8 F 2.3
4:00 166.5 F 5.4
6:00 187.9 F 9.4
8:00 205.2 F 10.9
10:00 216.0 F 7.0
12:00 228.4 F 11.6
14:00 231.4 F 8.7
16:00 240.1 F 11.8
18:00 244.1 F 13.1
20:00 255.3 F 19.0
22:00 265.8 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.9 F 29.8
2:00 271.5 F 29.0
3:00 272.2 F 29.8
4:00 272.6 F 29.5

Min 271.2 F 29.0
Max 273.1 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.7 F 0.0
5:00 230.0 F 0.0
10:00 229.6 F 0.0
15:00 231.1 F 0.0
20:00 228.1 F 0.0
25:00 236.8 F 0.0
30:00 224.3 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 23:52
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:02:43
CYCLE COMPLETE

06/17/2021 11:39 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____