
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001533
15:36 23/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:36
STERILIZING 3:34
136.3 C 226kPa 3:34
136.2 C 224kPa 4:04
136.4 C 223kPa 4:34
136.2 C 224kPa 5:04
136.4 C 226kPa 5:34
136.4 C 224kPa 6:04
136.4 C 224kPa 6:34
136.4 C 226kPa 7:04
136.4 C 226kPa 7:04
VENTING 7:04
AIR DRYING 8:01
CYCLE COMPLETE 8:05

Digital Signature #
B1D2CFFB83DF586B
