
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 002471
17:39 28/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:20
PRESSURIZING 2:25
STERILIZING 3:32
136.4 C 225kPa 3:32
136.1 C 224kPa 4:02
136.2 C 223kPa 4:32
136.3 C 225kPa 5:02
136.4 C 227kPa 5:32
136.4 C 227kPa 6:02
136.2 C 224kPa 6:32
136.4 C 227kPa 7:02
136.4 C 227kPa 7:02
VENTING 7:02
AIR DRYING 7:56
CYCLE COMPLETE 8:14

Digital Signature #
A97266071A041F5C
