

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2889

Name: RITTER 1  
Sterilizer ID: V2178198  
02/13/2023 10:03 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 180.3 F 0.0  
2:00 173.9 F 3.3  
4:00 181.0 F 7.3  
6:00 200.7 F 11.9  
8:00 215.2 F 5.3  
10:00 229.2 F 10.2  
12:00 231.8 F 7.8  
14:00 242.0 F 12.3  
16:00 239.8 F 10.3  
18:00 252.3 F 16.6  
20:00 263.4 F 23.0  
22:00 272.2 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.6 F 29.2  
2:00 271.8 F 28.4  
3:00 272.0 F 28.9  
4:00 272.5 F 29.1

Min 271.4 F 28.3  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.9 F 0.0  
5:00 229.3 F 0.0  
10:00 237.2 F 0.0  
15:00 229.7 F 0.0  
20:00 229.4 F 0.0  
25:00 228.0 F 0.0  
30:00 229.5 F 0.0  
Drying Complete

FILLING: 1:06  
HEATING: 22:14  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:00:35  
CYCLE COMPLETE

02/13/2023 11:03 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_