
Midmark M11 - v1.0.0
Total Cycles: 3370

Name: RITTER 1
Sterilizer ID: V2178198
06/12/2023 07:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.4 F 0.0
2:00 77.2 F 0.0
4:00 111.9 F 0.0
6:00 128.9 F 0.0
8:00 158.5 F 0.6
10:00 177.4 F 3.1
12:00 190.2 F 5.8
14:00 203.4 F 9.0
16:00 221.0 F 12.3
18:00 222.9 F 5.5
20:00 232.5 F 9.3
22:00 241.6 F 13.5
24:00 238.6 F 10.4
26:00 237.8 F 8.5
28:00 248.4 F 14.4
30:00 259.4 F 20.3
32:00 269.4 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.7 F 28.7
2:00 271.3 F 28.0
3:00 272.6 F 28.8
4:00 271.6 F 28.1

Min 271.2 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.9 F 0.0
5:00 203.0 F 0.0
10:00 197.7 F 0.0
15:00 205.9 F 0.0
20:00 207.5 F 0.0
25:00 206.6 F 0.0
30:00 203.2 F 0.0

Drying Complete

FILLING: 0:46
HEATING: 33:03
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:11:16
CYCLE COMPLETE

06/12/2023 09:10 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____