
Midmark M11 - v1.0.0
Total Cycles: 764

Name: RITTER 1
Sterilizer ID: V2178198
04/01/2021 08:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.4 F 0.0
2:00 77.7 F 0.0
4:00 80.2 F 0.0
6:00 101.6 F 0.1
8:00 135.4 F 0.2
10:00 177.5 F 1.4
12:00 202.6 F 4.8
14:00 210.2 F 8.3
16:00 222.3 F 11.9
18:00 228.6 F 6.8
20:00 237.1 F 11.9
22:00 236.8 F 9.1
24:00 238.3 F 9.3
26:00 248.3 F 14.2
28:00 260.4 F 20.9
30:00 271.2 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 271.8 F 28.3
2:00 272.5 F 28.8
3:00 271.7 F 28.5
4:00 271.6 F 28.2

Min 271.4 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.1 F 0.0
5:00 222.7 F 0.0
10:00 239.8 F 0.0
15:00 228.5 F 0.0
20:00 221.2 F 0.0
25:00 224.2 F 0.0
30:00 223.4 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 30:37
STERILIZING: 4:00
VENTING: 2:59
DRYING: 30:00
TOTAL CYCLE: 01:08:34
CYCLE COMPLETE

04/01/2021 09:22 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____