
Midmark M11 - v1.0.5
Total Cycles: 3595

Name: RITTER2
Sterilizer ID: V2313228
05/07/2024 10:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 114.9 F 0.0
2:00 119.0 F 0.4
4:00 117.0 F 2.1
6:00 144.0 F 4.7
8:00 177.0 F 8.0
10:00 184.5 F 11.9
12:00 208.9 F 4.6
14:00 219.0 F 8.4
16:00 229.8 F 12.6
18:00 231.3 F 8.8
20:00 241.2 F 12.4
22:00 241.6 F 12.0
24:00 252.7 F 17.6
26:00 262.6 F 23.5
28:00 272.1 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.4
1:00 272.3 F 30.2
2:00 272.4 F 29.9
3:00 271.7 F 29.5
4:00 271.2 F 29.4

Min 271.2 F 29.2
Max 272.7 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.6 F 0.0
5:00 209.4 F 0.0
10:00 219.6 F 0.0
15:00 230.5 F 0.0
20:00 228.0 F 0.0
25:00 226.0 F 0.0
30:00 228.4 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 28:37
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:07:07
CYCLE COMPLETE

05/07/2024 11:36 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____