
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002218
11:09 05/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:57
PRESSURIZING 2:03
STERILIZING 3:25
136.4 C 227kPa 3:25
136.3 C 224kPa 3:55
136.0 C 223kPa 4:25
136.3 C 225kPa 4:55
136.3 C 225kPa 5:25
136.1 C 223kPa 5:55
136.3 C 225kPa 6:25
136.4 C 226kPa 6:55
136.4 C 226kPa 6:55
VENTING 6:55
AIR DRYING 7:54
CYCLE COMPLETE 10:10

Digital Signature #
6E05BFD44D563F04
