
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001368
15:08 28/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:48
PRESSURIZING 3:50
STERILIZING 5:05
136.0 C 221kPa 5:05
135.9 C 222kPa 5:35
135.9 C 221kPa 6:05
136.5 C 226kPa 6:35
136.4 C 225kPa 7:05
136.2 C 224kPa 7:35
136.4 C 225kPa 8:05
136.2 C 223kPa 8:35
136.2 C 223kPa 8:35
VENTING 8:35
AIR DRYING 9:33
CYCLE COMPLETE 9:35

Digital Signature #
B348E81921474021
