
Midmark M11 - v1.0.0
Total Cycles: 732

Name: RITTER 1
Sterilizer ID: V2178198
03/18/2021 12:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.0 F 0.0
2:00 157.5 F 2.7
4:00 178.5 F 6.8
6:00 204.3 F 12.4
8:00 218.2 F 7.3
10:00 232.0 F 13.2
12:00 239.0 F 10.5
14:00 235.9 F 8.7
16:00 251.3 F 15.9
18:00 264.0 F 23.2
20:00 270.9 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 272.5 F 28.9
1:00 272.6 F 29.2
2:00 272.6 F 29.5
3:00 271.8 F 29.0
4:00 272.3 F 29.0

Min 271.6 F 28.8
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.1 F 0.0
5:00 232.1 F 0.0
10:00 227.4 F 0.0
15:00 227.3 F 0.0
20:00 225.3 F 0.0
25:00 222.1 F 0.0
30:00 223.4 F 0.0
Drying Complete

FILLING: 1:24
HEATING: 20:31
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 00:58:56
CYCLE COMPLETE

03/18/2021 01:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____