
Midmark M11 - v1.0.0
Total Cycles: 1605

Name: RITTER 1
Sterilizer ID: V2178198
02/08/2022 08:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.6 F 0.0
2:00 185.4 F 2.9
4:00 186.7 F 7.3
6:00 208.5 F 12.7
8:00 225.0 F 7.3
10:00 238.2 F 12.8
12:00 237.8 F 10.2
14:00 237.1 F 8.0
16:00 249.9 F 15.2
18:00 261.9 F 22.0
20:00 272.7 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.5 F 28.9
2:00 271.5 F 28.3
3:00 272.6 F 29.0
4:00 271.7 F 28.2

Min 271.4 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.0 F 0.0
5:00 226.7 F 0.0
10:00 231.9 F 0.0
15:00 228.7 F 0.0
20:00 227.7 F 0.0
25:00 229.8 F 0.0
30:00 228.7 F 0.0
Drying Complete

FILLING: 1:29
HEATING: 20:24
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:07
CYCLE COMPLETE

02/08/2022 09:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____