
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001506
12:21 06/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:56
PRESSURIZING 2:03
STERILIZING 2:56
136.0 C 220kPa 2:56
136.3 C 220kPa 3:26
136.4 C 221kPa 3:56
135.9 C 221kPa 4:26
136.2 C 224kPa 4:56
136.1 C 222kPa 5:26
136.2 C 223kPa 5:56
136.4 C 224kPa 6:26
136.4 C 224kPa 6:26
VENTING 6:26
AIR DRYING 7:14
CYCLE COMPLETE 37:14

Digital Signature #
2D608C6E6AC2FF9D
