

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5171

Name: RITTER 1  
Sterilizer ID: V2178198  
12/06/2024 05:31 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 144.5 F 0.0  
2:00 141.6 F 0.1  
4:00 175.8 F 0.5  
6:00 199.4 F 2.1  
8:00 222.2 F 7.1  
10:00 228.9 F 12.1  
12:00 232.9 F 8.1  
14:00 242.7 F 11.9  
16:00 243.8 F 12.1  
18:00 241.5 F 10.9  
20:00 255.6 F 18.0  
22:00 267.7 F 25.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.0  
1:00 272.3 F 28.5  
2:00 271.8 F 28.3  
3:00 272.3 F 28.5  
4:00 271.6 F 28.2

Min 271.4 F 27.9  
Max 272.6 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.9 F 0.0  
5:00 196.9 F 0.0  
10:00 195.8 F 0.0  
15:00 194.4 F 0.0  
20:00 195.0 F 0.0  
25:00 194.2 F 0.0  
30:00 194.7 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 23:18  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 01:01:28  
CYCLE COMPLETE

12/06/2024 06:32 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_