
Midmark M11 - v1.0.5
Total Cycles: 4456

Name: RITTER2
Sterilizer ID: V2313228
03/25/2025 05:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.3 F 0.0
2:00 170.4 F 3.1
4:00 193.5 F 8.0
6:00 215.3 F 13.5
8:00 225.3 F 8.9
10:00 235.3 F 10.5
12:00 240.8 F 12.4
14:00 241.2 F 11.5
16:00 254.3 F 18.3
18:00 266.0 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 29.2
1:00 272.6 F 29.9
2:00 271.7 F 29.3
3:00 272.4 F 29.6
4:00 271.8 F 29.6

Min 271.6 F 29.1
Max 272.6 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.2 F 0.0
5:00 196.1 F 0.0
10:00 209.5 F 0.0
15:00 205.3 F 0.0
20:00 207.6 F 0.0
25:00 213.4 F 0.0
30:00 212.6 F 0.0
Drying Complete

FILLING: 2:06
HEATING: 19:29
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 00:59:07
CYCLE COMPLETE

03/25/2025 06:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____