
Midmark M11 - v1.0.0
Total Cycles: 2996

Name: RITTER 1
Sterilizer ID: V2178198
03/13/2023 12:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 137.3 F 0.0
2:00 140.8 F 0.6
4:00 164.2 F 2.4
6:00 196.9 F 5.7
8:00 221.4 F 10.5
10:00 225.3 F 5.6
12:00 236.3 F 10.4
14:00 235.3 F 7.6
16:00 245.1 F 13.4
18:00 243.6 F 12.0
20:00 256.3 F 18.7
22:00 267.5 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.5 F 29.0
2:00 271.3 F 28.2
3:00 272.2 F 29.1
4:00 272.1 F 28.6

Min 271.2 F 28.2
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.7 F 0.0
5:00 229.3 F 0.0
10:00 234.0 F 0.0
15:00 222.2 F 0.0
20:00 221.9 F 0.0
25:00 219.5 F 0.0
30:00 220.8 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 23:22
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:01:24
CYCLE COMPLETE

03/13/2023 01:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____