

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1803

Name: RITTER 1  
Sterilizer ID: V2178198  
04/07/2022 09:03 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 121.4 F 0.0  
2:00 120.6 F 0.0  
4:00 135.1 F 1.7  
6:00 166.3 F 5.1  
8:00 188.6 F 8.9  
10:00 205.4 F 13.1  
12:00 221.3 F 6.4  
14:00 232.0 F 11.1  
16:00 230.5 F 7.2  
18:00 243.8 F 13.0  
20:00 241.8 F 11.3  
22:00 254.6 F 17.7  
24:00 265.6 F 24.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 272.5 F 29.0  
2:00 271.4 F 28.3  
3:00 272.7 F 29.2  
4:00 272.1 F 28.5

Min 271.4 F 28.3  
Max 272.9 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.9 F 0.0  
5:00 216.8 F 0.0  
10:00 233.8 F 0.0  
15:00 203.5 F 0.0  
20:00 233.2 F 0.0  
25:00 224.6 F 0.0  
30:00 236.7 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 25:46  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:03:50  
CYCLE COMPLETE

04/07/2022 10:06 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_