
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000034
15:13 24/01/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CYCLE ABORTED 3:16
CYCLE FAULT 15
NOT STERILE
END TIME 3:16

Digital Signature #
8A99702F4BA37576
