
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001071
9:04 11/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:44
PRESSURIZING 3:44
STERILIZING 4:46
136.5 C 224kPa 4:46
136.3 C 224kPa 5:16
CYCLE ABORTED 5:23
STOP BUTTON PRESSED
NOT STERILE
END TIME 5:23

Digital Signature #
001F5C6D2849BC02
