

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000935  
10:07 17/02/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:49  
PRESSURIZING 3:52  
STERILIZING 5:17  
136.0 C 221kPa 5:17  
136.4 C 233kPa 5:47  
136.4 C 231kPa 6:17  
136.4 C 230kPa 6:47  
136.4 C 232kPa 7:17  
136.2 C 230kPa 7:47  
136.4 C 232kPa 8:17  
136.3 C 231kPa 8:47  
136.3 C 231kPa 8:47  
VENTING 8:47  
AIR DRYING 9:44  
CYCLE COMPLETE 24:44  
-----

Digital Signature #  
5B58600A352FC74F  
-----