

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2043

Name: RITTER2  
Sterilizer ID: V2313228  
02/13/2023 04:47 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 136.9 F 0.0  
2:00 137.7 F 0.0  
4:00 167.9 F 1.7  
6:00 193.0 F 6.2  
8:00 212.7 F 11.5  
10:00 223.5 F 6.9  
12:00 237.6 F 12.5  
14:00 238.5 F 10.7  
16:00 237.6 F 9.7  
18:00 252.8 F 17.1  
20:00 265.2 F 24.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.8  
1:00 272.4 F 29.2  
2:00 272.5 F 29.6  
3:00 271.9 F 28.8  
4:00 272.9 F 29.6

Min 271.5 F 28.6  
Max 272.9 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 174.5 F 0.0  
5:00 213.3 F 0.0  
10:00 232.7 F 0.0  
15:00 229.4 F 0.0  
20:00 219.9 F 0.0  
25:00 219.3 F 0.0  
30:00 219.4 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 21:39  
STERILIZING: 4:00  
VENTING: 3:02  
DRYING: 30:00  
TOTAL CYCLE: 00:59:40  
CYCLE COMPLETE

02/13/2023 05:46 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_