
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001980
13:32 26/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:12
PRESSURIZING 2:16
STERILIZING 3:23
136.5 C 225kPa 3:23
136.1 C 223kPa 3:53
136.4 C 226kPa 4:23
136.2 C 224kPa 4:53
136.2 C 224kPa 5:23
136.3 C 225kPa 5:53
136.2 C 224kPa 6:23
136.4 C 226kPa 6:53
136.4 C 226kPa 6:53
VENTING 6:53
AIR DRYING 7:46
CYCLE COMPLETE 8:23

Digital Signature #
77667C5B1FF817A6
