
Midmark M11 - v1.0.5
Total Cycles: 2435

Name: RITTER2
Sterilizer ID: V2313228
05/30/2023 11:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.6 F 0.0
2:00 140.6 F 0.0
4:00 163.3 F 2.7
6:00 188.1 F 7.3
8:00 211.8 F 12.5
10:00 222.9 F 7.3
12:00 236.3 F 12.4
14:00 236.1 F 9.9
16:00 237.4 F 8.8
18:00 249.4 F 15.4
20:00 261.5 F 22.4
22:00 272.5 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.7 F 29.8
2:00 271.8 F 29.0
3:00 271.9 F 29.6
4:00 272.6 F 29.7

Min 271.3 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.6 F 0.0
5:00 216.5 F 0.0
10:00 226.4 F 0.0
15:00 228.7 F 0.0
20:00 228.5 F 0.0
25:00 223.9 F 0.0
30:00 222.8 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 22:27
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:32
CYCLE COMPLETE

05/30/2023 12:46 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____