

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 003534  
14:32 26/04/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:37  
PRESSURIZING 2:41  
STERILIZING 3:55  
135.9 C 220kPa 3:55  
136.4 C 222kPa 4:25  
136.4 C 223kPa 4:55  
136.4 C 222kPa 5:25  
136.4 C 225kPa 5:55  
136.2 C 222kPa 6:25  
136.1 C 218kPa 6:55  
136.2 C 220kPa 7:25  
136.2 C 220kPa 7:25  
VENTING 7:25  
AIR DRYING 8:20  
CYCLE COMPLETE 9:00

-----  
Digital Signature #  
2E81F5C71178E804  
-----