

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3406

Name: RITTER2  
Sterilizer ID: V2313228  
02/23/2024 10:47 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 130.6 F 0.0  
2:00 130.6 F 0.2  
4:00 133.4 F 1.8  
6:00 169.9 F 4.9  
8:00 185.5 F 8.8  
10:00 203.5 F 13.0  
12:00 219.5 F 7.3  
14:00 234.3 F 11.9  
16:00 231.5 F 8.8  
18:00 239.1 F 11.1  
20:00 243.9 F 13.1  
22:00 255.3 F 19.0  
24:00 265.6 F 25.5  
26:00 272.8 F 30.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.7  
1:00 271.7 F 29.9  
2:00 271.5 F 30.0  
3:00 271.6 F 30.0  
4:00 271.5 F 29.8

Min 271.4 F 29.6  
Max 272.6 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.9 F 0.0  
5:00 226.1 F 0.0  
10:00 195.3 F 0.0  
15:00 191.1 F 0.0  
20:00 188.6 F 0.0  
25:00 188.4 F 0.0  
30:00 189.4 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 26:20  
STERILIZING: 4:00  
VENTING: 3:37  
DRYING: 30:00  
TOTAL CYCLE: 01:04:50  
CYCLE COMPLETE

02/23/2024 11:51 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_