

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2856

Name: RITTER2  
Sterilizer ID: V2313228  
09/08/2023 03:59 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 178.1 F 0.0  
2:00 178.3 F 3.4  
4:00 197.1 F 8.1  
6:00 213.3 F 13.1  
8:00 226.1 F 8.3  
10:00 237.3 F 13.5  
12:00 238.9 F 11.4  
14:00 238.6 F 10.3  
16:00 251.8 F 16.9  
18:00 263.1 F 23.5  
20:00 272.6 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.2  
1:00 272.7 F 29.8  
2:00 271.5 F 29.0  
3:00 272.3 F 30.0  
4:00 272.0 F 29.4

Min 271.2 F 29.0  
Max 272.9 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.4 F 0.0  
5:00 209.7 F 0.0  
10:00 215.6 F 0.0  
15:00 217.8 F 0.0  
20:00 212.0 F 0.0  
25:00 214.0 F 0.0  
30:00 212.8 F 0.0  
Drying Complete

FILLING: 2:14  
HEATING: 20:16  
STERILIZING: 4:00  
VENTING: 3:35  
DRYING: 30:00  
TOTAL CYCLE: 01:00:04  
CYCLE COMPLETE

09/08/2023 04:59 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_