
Midmark M11 - v1.0.0
Total Cycles: 4958

Name: RITTER 1
Sterilizer ID: V2178198
10/03/2024 10:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.0 F 0.0
2:00 170.5 F 0.8
4:00 173.2 F 3.5
6:00 196.3 F 7.4
8:00 210.6 F 12.0
10:00 221.0 F 5.6
12:00 230.4 F 10.3
14:00 233.1 F 8.0
16:00 242.6 F 12.5
18:00 240.6 F 10.7
20:00 252.7 F 16.8
22:00 263.4 F 23.1
24:00 272.3 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.4 F 29.1
2:00 271.7 F 28.4
3:00 271.8 F 28.9
4:00 272.2 F 29.0

Min 271.3 F 28.2
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.4 F 0.0
5:00 202.4 F 0.0
10:00 207.6 F 0.0
15:00 224.0 F 0.0
20:00 230.4 F 0.0
25:00 231.0 F 0.0
30:00 230.4 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 24:14
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:03:11
CYCLE COMPLETE

10/03/2024 11:06 AM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____