
Midmark M11 - v1.0.0
Total Cycles: 3727

Name: RITTER 1
Sterilizer ID: V2178198
09/06/2023 02:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 124.2 F 0.0
2:00 127.3 F 0.0
4:00 152.3 F 1.2
6:00 177.5 F 4.0
8:00 200.9 F 7.9
10:00 216.4 F 12.4
12:00 226.7 F 6.9
14:00 238.7 F 11.9
16:00 235.8 F 8.9
18:00 239.7 F 10.1
20:00 247.3 F 13.7
22:00 259.5 F 20.4
24:00 270.5 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.0 F 28.4
2:00 272.1 F 28.8
3:00 271.7 F 28.1
4:00 272.4 F 28.9

Min 271.2 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.2 F 0.0
5:00 225.8 F 0.0
10:00 219.3 F 0.0
15:00 228.2 F 0.0
20:00 228.7 F 0.0
25:00 223.3 F 0.0
30:00 225.5 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 24:46
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:03:02
CYCLE COMPLETE

09/06/2023 03:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____