

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1984

Name: RITTER 1  
Sterilizer ID: V2178198  
06/01/2022 07:08 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.7 F 0.0  
2:00 86.2 F 0.1  
4:00 104.8 F 1.2  
6:00 119.0 F 3.5  
8:00 147.9 F 6.5  
10:00 167.3 F 9.9  
12:00 203.7 F 8.8  
14:00 207.5 F 5.6  
16:00 219.6 F 9.1  
18:00 230.4 F 12.9  
20:00 230.6 F 8.2  
22:00 241.6 F 13.0  
24:00 239.3 F 10.4  
26:00 251.5 F 16.4  
28:00 262.2 F 22.5  
30:00 272.2 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.7  
1:00 272.5 F 29.4  
2:00 271.6 F 28.6  
3:00 272.2 F 29.5  
4:00 272.1 F 29.0

Min 271.3 F 28.6  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.9 F 0.0  
5:00 204.8 F 0.0  
10:00 223.3 F 0.0  
15:00 230.2 F 0.0  
20:00 233.7 F 0.0  
25:00 233.4 F 0.0  
30:00 225.8 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 30:31  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:08:25  
CYCLE COMPLETE

06/01/2022 08:16 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_