
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 005255
11:46 20/03/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:37
STERILIZING 3:39
136.4 C 220kPa 3:39
136.3 C 223kPa 4:09
136.2 C 223kPa 4:39
136.4 C 225kPa 5:09
136.4 C 223kPa 5:39
136.2 C 224kPa 6:09
136.4 C 225kPa 6:39
136.4 C 223kPa 7:09
136.4 C 223kPa 7:09
VENTING 7:09
AIR DRYING 8:10
CYCLE COMPLETE 11:56

Digital Signature #
1103BBBA3394C5F2
