
Midmark M11 - v1.0.5
Total Cycles: 3933

Name: RITTER2
Sterilizer ID: V2313228
09/11/2024 06:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.6 F 0.0
2:00 155.5 F 2.1
4:00 181.2 F 5.9
6:00 212.2 F 10.9
8:00 221.2 F 5.6
10:00 231.9 F 10.7
12:00 231.4 F 7.6
14:00 242.3 F 11.9
16:00 244.8 F 13.0
18:00 257.1 F 19.5
20:00 268.2 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.2
1:00 272.5 F 29.4
2:00 272.8 F 29.8
3:00 271.7 F 29.3
4:00 272.7 F 29.5

Min 271.6 F 29.1
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.4 F 0.0
5:00 234.7 F 0.0
10:00 225.3 F 0.0
15:00 225.5 F 0.0
20:00 218.3 F 0.0
25:00 204.4 F 0.0
30:00 185.6 F 0.0
Drying Complete

FILLING: 1:54
HEATING: 21:56
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:01:16
CYCLE COMPLETE

09/11/2024 07:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____