
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004159
10:21 08/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
PRESSURIZING 2:43
STERILIZING 3:50
136.0 C 220kPa 3:50
136.0 C 223kPa 4:20
136.3 C 223kPa 4:50
136.3 C 222kPa 5:20
136.0 C 222kPa 5:50
136.3 C 224kPa 6:20
136.4 C 224kPa 6:50
136.4 C 224kPa 7:20
136.4 C 224kPa 7:20
VENTING 7:20
AIR DRYING 8:13
CYCLE COMPLETE 8:21

Digital Signature #
185EB774D1D21B51
