
Midmark M11 - v1.0.0
Total Cycles: 2450

Name: RITTER 1
Sterilizer ID: V2178198
10/18/2022 12:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 208.3 F 0.0
2:00 201.6 F 4.3
4:00 223.0 F 10.3
6:00 227.0 F 6.2
8:00 238.8 F 12.2
10:00 237.6 F 10.1
12:00 238.0 F 8.6
14:00 251.2 F 16.1
16:00 263.7 F 23.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.5 F 29.1
2:00 271.8 F 28.8
3:00 272.5 F 29.1
4:00 271.6 F 28.5

Min 271.6 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.8 F 0.0
5:00 229.9 F 0.0
10:00 235.2 F 0.0
15:00 228.3 F 0.0
20:00 231.6 F 0.0
25:00 219.9 F 0.0
30:00 222.7 F 0.0

Drying Complete

FILLING: 3:30
HEATING: 17:58
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:44
CYCLE COMPLETE

10/18/2022 01:50 PM
ACCEPTED: X
REJECTED:
Operator: U02