

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1355

Name: RITTER2  
Sterilizer ID: V2313228  
07/27/2022 10:59 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 182.5 F 0.0  
2:00 179.5 F 0.1  
4:00 185.5 F 2.4  
6:00 206.9 F 7.1  
8:00 225.4 F 12.3  
10:00 230.2 F 8.2  
12:00 241.7 F 13.6  
14:00 240.7 F 11.7  
16:00 239.6 F 10.6  
18:00 252.9 F 17.2  
20:00 264.1 F 24.0  
22:00 271.7 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.9  
1:00 272.8 F 29.8  
2:00 272.4 F 29.3  
3:00 271.5 F 28.9  
4:00 272.4 F 29.7

Min 271.4 F 28.8  
Max 272.9 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.7 F 0.0  
5:00 219.7 F 0.0  
10:00 227.3 F 0.0  
15:00 232.8 F 0.0  
20:00 228.4 F 0.0  
25:00 228.4 F 0.0  
30:00 226.0 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 22:07  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:00:17  
CYCLE COMPLETE

07/27/2022 11:59 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_