
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.7uS / 0.4ppm
CYCLE NUMBER 001446
13:47 07/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:56
PRESSURIZING 2:00
STERILIZING 3:04
136.4 C 224kPa 3:04
136.4 C 226kPa 3:34
136.4 C 226kPa 4:04
136.2 C 224kPa 4:34
136.4 C 226kPa 5:04
136.4 C 226kPa 5:34
136.4 C 225kPa 6:04
136.4 C 226kPa 6:34
136.4 C 226kPa 6:34
VENTING 6:34
AIR DRYING 7:27
CYCLE COMPLETE 22:27

Digital Signature #
6CD79C97F772C3DB
