
Midmark M11 - v1.0.5
Total Cycles: 3976

Name: RITTER2
Sterilizer ID: V2313228
09/26/2024 08:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.3 F 0.0
2:00 70.7 F 0.0
4:00 76.4 F 0.0
6:00 102.4 F 0.1
8:00 143.4 F 0.3
10:00 170.8 F 0.9
12:00 190.7 F 4.0
14:00 204.6 F 7.5
16:00 218.8 F 11.3
18:00 222.0 F 6.1
20:00 232.6 F 10.3
22:00 234.6 F 8.7
24:00 243.3 F 12.8
26:00 241.8 F 11.5
28:00 254.0 F 17.8
30:00 264.9 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.6 F 29.4
2:00 271.4 F 28.8
3:00 272.6 F 29.5
4:00 271.6 F 28.7

Min 271.4 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.6 F 0.0
5:00 193.3 F 0.0
10:00 198.1 F 0.0
15:00 191.1 F 0.0
20:00 190.4 F 0.0
25:00 188.6 F 0.0
30:00 191.7 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 31:57
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:10:07
CYCLE COMPLETE

09/26/2024 09:34 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____