
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001389
14:53 16/12/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:05
PRESSURIZING 3:09
STERILIZING 4:20
136.2 C 223kPa 4:20
136.4 C 225kPa 4:50
136.3 C 224kPa 5:20
136.2 C 223kPa 5:50
136.2 C 223kPa 6:20
136.2 C 225kPa 6:50
136.4 C 224kPa 7:20
136.4 C 226kPa 7:50
136.4 C 226kPa 7:50
VENTING 7:50
AIR DRYING 8:46
CYCLE COMPLETE 23:46

Digital Signature #
4086E9A7370C57C4
