
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 007986
13:40 21/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:35
PRESSURIZING 3:36
STERILIZING 4:35
136.4 C 223kPa 4:35
136.4 C 224kPa 5:05
136.4 C 223kPa 5:35
136.3 C 222kPa 6:05
136.3 C 223kPa 6:35
136.3 C 223kPa 7:05
136.1 C 220kPa 7:35
136.4 C 220kPa 8:05
136.4 C 220kPa 8:05
VENTING 8:05
AIR DRYING 9:00
CYCLE COMPLETE 9:09

Digital Signature #
97BB18B316D70F70
