

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2570

Name: RITTER 1  
Sterilizer ID: V2178198  
11/16/2022 03:58 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 197.5 F 0.0  
2:00 192.0 F 5.1  
4:00 202.8 F 10.2  
6:00 219.1 F 5.1  
8:00 231.8 F 10.5  
10:00 231.5 F 7.5  
12:00 241.9 F 11.4  
14:00 245.9 F 13.3  
16:00 259.1 F 20.4  
18:00 270.8 F 27.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.5  
1:00 272.2 F 29.3  
2:00 272.0 F 28.8  
3:00 271.6 F 28.7  
4:00 272.6 F 29.3  
  
Min 271.3 F 28.4  
Max 272.6 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.1 F 0.0  
5:00 232.1 F 0.0  
10:00 235.2 F 0.0  
15:00 228.3 F 0.0  
20:00 226.1 F 0.0  
25:00 231.5 F 0.0  
30:00 227.8 F 0.0  
  
Drying Complete

FILLING: 1:38  
HEATING: 18:41  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:57:32  
CYCLE COMPLETE

11/16/2022 04:55 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_