
Midmark M11 - v1.0.5
Total Cycles: 2244

Name: RITTER2
Sterilizer ID: V2313228
04/13/2023 08:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 112.7 F 0.0
2:00 114.9 F 0.0
4:00 117.8 F 0.2
6:00 156.7 F 3.5
8:00 185.4 F 7.1
10:00 205.2 F 11.4
12:00 217.5 F 5.6
14:00 228.1 F 10.2
16:00 233.3 F 8.5
18:00 241.2 F 12.4
20:00 239.8 F 10.9
22:00 252.1 F 17.0
24:00 263.2 F 23.5
26:00 272.4 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.1
1:00 272.5 F 29.8
2:00 271.3 F 29.0
3:00 272.6 F 30.1
4:00 272.0 F 29.9

Min 270.9 F 28.9
Max 273.3 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.7 F 0.0
5:00 204.8 F 0.0
10:00 206.7 F 0.0
15:00 207.1 F 0.0
20:00 212.4 F 0.0
25:00 215.2 F 0.0
30:00 215.4 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 26:16
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:04:23
CYCLE COMPLETE

04/13/2023 10:02 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____