

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5416

Name: RITTER 1  
Sterilizer ID: V2178198  
03/14/2025 10:34 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 112.6 F 0.0  
2:00 116.9 F 1.1  
4:00 129.9 F 3.4  
6:00 152.0 F 6.4  
8:00 175.8 F 9.8  
10:00 195.9 F 13.4  
12:00 205.0 F 3.9  
14:00 216.1 F 7.4  
16:00 227.4 F 11.1  
18:00 231.8 F 6.8  
20:00 241.3 F 11.5  
22:00 236.2 F 8.8  
24:00 248.3 F 14.4  
26:00 258.6 F 20.0  
28:00 267.6 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.7  
1:00 272.2 F 29.3  
2:00 272.9 F 29.3  
3:00 272.2 F 28.9  
4:00 271.4 F 28.6

Min 271.3 F 28.5  
Max 273.0 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.4 F 0.0  
5:00 190.6 F 0.0  
10:00 182.7 F 0.0  
15:00 182.1 F 0.0  
20:00 184.4 F 0.0  
25:00 190.7 F 0.0  
30:00 195.3 F 0.0  
Drying Complete

FILLING: 0:45  
HEATING: 29:36  
STERILIZING: 4:00  
VENTING: 3:55  
DRYING: 30:00  
TOTAL CYCLE: 01:08:16  
CYCLE COMPLETE

03/14/2025 11:42 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_