
Midmark M11 - v1.0.0
Total Cycles: 3559

Name: RITTER 1
Sterilizer ID: V2178198
07/21/2023 07:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.9 F 0.0
2:00 86.9 F 0.0
4:00 91.0 F 0.0
6:00 120.4 F 0.0
8:00 150.6 F 1.6
10:00 174.8 F 3.4
12:00 187.8 F 6.0
14:00 220.8 F 9.3
16:00 222.2 F 12.7
18:00 221.8 F 5.5
20:00 231.9 F 9.2
22:00 241.3 F 13.2
24:00 237.7 F 9.9
26:00 237.9 F 9.2
28:00 247.0 F 13.7
30:00 258.0 F 19.6
32:00 268.0 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.7 F 28.8
2:00 271.5 F 28.1
3:00 272.7 F 29.0
4:00 271.8 F 28.2

Min 271.4 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.1 F 0.0
5:00 213.5 F 0.0
10:00 195.5 F 0.0
15:00 193.5 F 0.0
20:00 198.5 F 0.0
25:00 203.3 F 0.0
30:00 194.5 F 0.0

Drying Complete

FILLING: 0:47
HEATING: 33:23
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:11:27
CYCLE COMPLETE

07/21/2023 08:45 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____