

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 002574  
8:11 14/03/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:00  
PRESSURIZING 2:06  
STERILIZING 3:26  
136.4 C 225kPa 3:26  
136.3 C 223kPa 3:56  
136.2 C 221kPa 4:26  
136.4 C 225kPa 4:56  
136.2 C 223kPa 5:26  
136.3 C 224kPa 5:56  
136.3 C 224kPa 6:26  
136.5 C 225kPa 6:56  
136.5 C 225kPa 6:56  
VENTING 6:56  
AIR DRYING 7:53  
CYCLE COMPLETE 27:47  
-----

Digital Signature #  
6B9BACCAB5BD85A4  
-----