
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006546
7:50 02/11/2023

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:57
PRESSURIZING 3:00
STERILIZING 4:03
136.4 C 223kPa 4:03
136.3 C 222kPa 4:33
136.2 C 218kPa 5:03
136.0 C 220kPa 5:33
136.1 C 222kPa 6:03
136.2 C 220kPa 6:33
136.3 C 222kPa 7:03
136.4 C 222kPa 7:33
136.4 C 222kPa 7:33
VENTING 7:33
AIR DRYING 8:21
CYCLE COMPLETE 38:21

Digital Signature #
1AA1FCBB50BB2DFC
